
MightySat[™]Rx Fingertip Pulse Oximeter

Featuring Masimo SET® Measure-through Motion and Low Perfusion™ Pulse Oximetry

Measurements











> MASIMO SET® - PROVEN TECHNOLOGY IN HOSPITALS WORLDWIDE

- > In a study comparing the ability of three pulse oximetry technologies to detect hypoxic events, Masimo SET° demonstrated the highest sensitivity and specificity during induced conditions of motion and low perfusion¹
- > Masimo SET® pulse oximetry helps clinicians monitor more than 100 million patients a year²

> DESIGNED FOR PERFORMANCE



> FREE DOWNLOADABLE MASIMO PROFESSIONAL HEALTH APP*


Display parameter data to a compatible smart device with Bluetooth® LE enabled models. Measurements can also integrate into Apple's Health app.**



VIEW measurements on compatible smart device*



TREND measurements over time, view on graphical display, and share the captured data



Example of data trend over 6 minutes


* The app is downloadable from the App store™ for iOS devices or Google Play™ store for select Android devices. For an up-to-date list of compatible smart devices, see:

www.masimoprofessionalhealth.com

> COMPARISON AT A GLANCE

	Masimo MightySat Rx	Nonin Onyx II 9560
Key Product Features		
Parameters: SpO2 and PR featuring Masimo Signal Extraction Technology (SET*)	Yes	No
Parameter: Perfusion Index (PI)	Yes	No
Parameter: Pleth Variability Index (PVI)	Yes	No
Pleth Waveform	Yes	No
Rotating and adjustable color screen for real-time display of all measurements	Yes	No
App includes: high resolution pleth waveform, smart tone audible feature, and trending functionality	Yes	No
Customizable settings with use of touchpad	Yes	No
Flexible finger pad	Yes	No

> RESULTS IN TWO EASY STEPS





Obtain measurements





Kit includes: MightySat Rx device, lanyard, batteries, and carrying case



Unique Parameters and Indicators

Perfusion Index (PI)

The ratio of the pulsatile blood flow to the non-pulsatile blood in peripheral tissue used to measure peripheral perfusion.

Pleth Variability Index (PVI)

A measure of the dynamic changes in the perfusion index (PI) that occur during the respiratory cycle.

Signal IQ (SIQ)

Located under the plethysmographic waveform. The height of the vertical line provides an assessment of the confidence in the SpO2 value displayed.

Plethysmographic (Pleth) Waveform

Real-time pulse pressure waveform.





High resolution view of the pleth waveform when viewed on a smart device.

	MightySat Rx P/N 9709	MightySat Rx with Bluetooth LE P/N 9809	MightySat Rx with Bluetooth LE & PVI P/N 9909
Masimo SET® Measure-through Motion and Low Perfusion™ Technology	•	•	•
Measurements SpO2 (Oxygen Saturation)	•	•	•
PR (Pulse Rate)	•	•	•
PI (Perfusion Index)	•	•	•
PVI (Pleth Variability Index)			•
Display Pleth Waveform	•	•	•
3.3 cm Color Display	•	•	•
Rotational Screen	•	•	•
Bluetooth LE		•	•
Compatible with Masimo Professional Health App*		•	•
Four Year Limited Warranty	•	•	•

MightySat Rx Specifications

DISPLAY RANGE	PHYSICAL CHARACTERISTICS	
Functional Oxygen Saturation (SpO2) 0-100% Pulse Rate (PR) 25-240 bpm Perfusion Index (PI) 0.02-20% Pleth Variability Index (PVI) 0-100%	Weight with Battery ⁴	
ACCURACY – (ARMS)3	No alarms on this product	
ACCURACY - (ARMS)S 70-100% Sp02 Accuracy Range 70-100% No Motion 2% Motion 3% Low Perfusion 2% PR Accuracy Range 25-240 bpm No Motion 3 bpm Motion 5 bpm Low Perfusion 3 bpm PATIENT WEIGHT Adult and Pediatric Patients >30 kg (66 lbs)	CLASSIFICATION PER IEC 60601-1 EMC Classification	
MEASUREMENT RESOLUTION	BATTERY	
SpO2	Operating	
ENVIRONMENTAL	Radio Modes	
Operating Temperature	Communication Compliance JFCC ID: VKF-MSAT01A USA FCC parts 15.247 Canada IC: 7362A-MSAT01A RSS-210 RSS-210 Europe EN 300 328 EN 301 489-17	

¹ Shah N et al. J. Clin Anesth. 2012 Aug; 24(5):385-91. ² Masimo data on file ³ A_{RMS} (Accuracy Root Mean Square) values determined using the MightySat Rx with Masimo SET* Oximetry Technology in a clinical study under motion and no motion conditions. ⁴ Weight is dependent on batteries used. ⁵ Based upon 15 hours of operation with screen brightness set to 50% and a spot check of 30 seconds.

Caution: Federal (USA) law restricts this device to sale by or on the order of a physician. See instructions for use for full prescribing information, including indications, contraindications, warnings, and precautions.







^{*}The app is downloadable from the App store™ for iOS devices or Google Play™ store for select Android devices. For an up-to-date list of compatible smart devices, see: www.masimoprofessionalhealth.com

^{**}Apple is a registered trademark of Apple Inc. registered in the U.S. and other countries.

Bluetooth is a registered trademark owned by Bluetooth SIG, Inc.

IOS is a trademark or registered trademark of Cisco in the U.S. and other countries and is used under license. Android is a trademark of Google Inc.